

Mi-2 Autoantibodies

Order Name: **Mi-2 Auto Ab**
Test Number: 6905993
Revision Date: 12/12/2022

| TEST NAME | METHODOLOGY | LOINC CODE |
|---------------------|-----------------------------|------------|
| Mi-2 Autoantibodies | <u>RIPA gel radiography</u> | 18485-3 |

| SPECIMEN REQUIREMENTS | | | | |
|-----------------------|---|---------------|--------------------|-----------------------|
| Specimen | Specimen Volume (min) | Specimen Type | Specimen Container | Transport Environment |
| Preferred | 1 mL (0.3 mL) | Serum | Clot Activator SST | Refrigerated |
| Instructions | <p>Notes: 0.3 mL (Note: This volume Does NOT allow for repeat testing.)</p> <p>Specimen Type: Red-top tube or gel-barrier tube</p> <p>Specimen Storage: Refrigerate or FREEZE</p> <p>Specimen Collection: Separate serum from cells within one hour of collection. Transfer to a plastic transport tube before shipping. To avoid delays in turnaround time when requesting multiple test on frozen samples, PLEASE SUBMIT SEPARATE FROZEN SPECIMENS FOR EACH TEST REQUESTED.</p> <p>Special Instructions: Testing is performed at Esoterix Inc. ESECF# 237</p> <p>Specimen Stability: Ambient: 7 days, Refrigerated : 14 days, Frozen: 60 days</p> | | | |

| GENERAL INFORMATION | |
|---------------------|--|
| Expected TAT | 14 - 21 days |
| Notes | Labcorp Test Code: 520000 |
| CPT Code(s) | 86235 |
| Service Provided By |  labcorp Oklahoma, Inc. |