


## TIBC/% Saturation Iron Profile Analyzer

Order Name: **AN %SAT/TI**  
Test Number: 4501525  
Revision Date: 01/16/2024

TEST NAME	METHODOLOGY	LOINC CODE
Total Iron Binding Capacity		2500-7
Iron .....	Ferene .....	2498-4
% Saturation TIBC		2502-3
UIBC	Ferene .....	2501-5

SPECIMEN REQUIREMENTS				
Specimen	Specimen Volume (min)	Specimen Type	Specimen Container	Transport Environment
Preferred	1 mL (0.5 mL)	Serum	Clot Activator SST (Gold Top)	Refrigerated
Instructions	Specimen stability: Ambient 10 hours, Refrigerated 7 days, Frozen 1 year			

GENERAL INFORMATION	
Testing Schedule	Daily
Expected TAT	1-2 days
Notes	<p>For ongoing iron determinations on a patient. The sample should be obtained at approximately the same time each day.</p> <p>The concentration may vary 30% throughout the day.</p> <p>If the Iron Level is less than 10 ug/dL then Estimation Iron with Total Iron Binding Capacity is ordered and resulted at no charge.</p> <p><b>Order Code:</b> [0691091] .Iron with Total Iron Binding Capacity</p> <p><b>Result Code:</b> [4501101] Estimated TIBC (LOINC 2500-7)</p> <p><b>Result Code:</b> [4501151] Estimated % Saturation (LOINC 2502-3)</p>
CPT Code(s)	83550, 83540
Service Provided By	 <b>labcorp</b> Oklahoma, Inc.