

**Lupus Panel (SLE)**

Order Name: **LUPUS PNL**  
 Test Number: 5500380  
 Revision Date: 03/15/2019

TEST NAME	METHODOLOGY	LOINC CODE
Anti-Nuclear Antibody Screen (ANA) with Reflex to Titer	Indirect Fluorescent Antibody	8061-4
Complement C3 and C4	Turbidometric	
DNA Double-Stranded Antibody (anti-ds DNA)	Chemiluminescence Assay	32677-7
ENA Screen	Chemiluminescence Assay	
<b>Lupus Panel Interpretation</b>		

SPECIMEN REQUIREMENTS				
Specimen	Specimen Volume (min)	Specimen Type	Specimen Container	Transport Environment
Preferred	<b>6.5 mL (4)</b>	Serum	<b>Clot Activator SST</b>	<b>Frozen</b>
Instructions	For the best stability for all testing within this panel please process specimen in the following way:  <b>Allow specimen to clot completely at room temperature, then separate serum from cells within one hour of collection.</b> <b>Combined Stability: Room Temperature N/A, Refrigerated: 48hrs, Frozen 8 days.</b>  See the individual assays for their specific stability options.			
GENERAL INFORMATION				
Testing Schedule	ANA screen, C3/C4 are Mon-Fri, ENA, ds-DNA are Mon, Wed, Fri			
Expected TAT	Assay dependant			
Clinical Use	To assist in the diagnosis and monitoring of SLE.			
Notes				
CPT Code(s)	86235; 86225; 86038; 86160x2			
Service Provided By	 <b>labcorp</b> Oklahoma, Inc.			