

**Galectin-3**

 Order Name: **GALECTIN 3**

Test Number: 2015057

Revision Date: 12/12/2022

TEST NAME	METHODOLOGY	LOINC CODE
Galectin-3	Quantitative Enzyme Immunoassay	62419-7

**SPECIMEN REQUIREMENTS**

Specimen	Specimen Volume (min)	Specimen Type	Specimen Container	Transport Environment
Preferred	<b>0.8 mL (0.4)</b>	<b>Plasma</b>	<b>EDTA (Lavender Top)</b>	<b>Refrigerated</b>
Alternate 1	<b>0.8 mL (0.4)</b>	<b>Serum</b>	<b>Clot Activator (Red Top, No-Gel)</b>	<b>Refrigerated</b>
Instructions	<b>Notes:</b> 0.4 mL (Note: This volume Does NOT allow for repeat testing.) <b>Specimen Type:</b> Red-top tube, gel-barrier tube OR lavender-top (EDTA) tube <b>Specimen Storage:</b> Room temperature Stability: Room temperature up to 22 days Refrigerated up to 22 days Frozen up to 12 months Stable x9 <b>Specimen Collection:</b> Transfer separated serum or plasma to a plastic transport tube. <b>Specimen Stability:</b> Ambient: Not Available, Refrigerated : Not Available, Frozen: Not Available			

**GENERAL INFORMATION**

Expected TAT	4 - 6 days
Notes	Labcorp Test Code: 004110
CPT Code(s)	82777
Service Provided By	 <b>labcorp</b> Oklahoma, Inc.